

CERTIFICATE OF COMPLIANCE

Certificate Number 20160104-E468659
Report Reference E468659-20140919
Issue Date 2016-January-04

Issued to: SILEGO TECHNOLOGY
1515 Wyatt Dr
Santa Clara CA 95054-1524

**This is to certify that
representative samples of**

**COMPONENT - PROTECTORS, LOW-VOLTAGE SOLID-
STATE OVERCURRENT**

Component – Low voltage solid-state overcurrent
protectors, Models SLG55544, SLG55546, SLGC55544,
SLGC55544B, SLGC55544C and SLGC55546, maybe
followed by V or VTR.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:
Additional Information:

UL 2367 Solid-state Overcurrent Protectors
See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance
capabilities and are intended for use as components of complete equipment submitted for investigation rather
than for direct separate installation in the field. The final acceptance of the component is dependent upon its
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

